



## King's Research Portal

DOI:

[10.1186/2197-7364-2-S1-A29](https://doi.org/10.1186/2197-7364-2-S1-A29)

*Document Version*

Publisher's PDF, also known as Version of record

[Link to publication record in King's Research Portal](#)

*Citation for published version (APA):*

Merida, I., Costes, N., Heckemann, R., & Hammers, A. (2015). Pseudo-CT generation in brain MR-PET attenuation correction: comparison of several multi-atlas methods. *EJNMMI Physics*, 2(Suppl 1), A29. <https://doi.org/10.1186/2197-7364-2-S1-A29>

### Citing this paper

Please note that where the full-text provided on King's Research Portal is the Author Accepted Manuscript or Post-Print version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version for pagination, volume/issue, and date of publication details. And where the final published version is provided on the Research Portal, if citing you are again advised to check the publisher's website for any subsequent corrections.

### General rights

Copyright and moral rights for the publications made accessible in the Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognize and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the Research Portal

### Take down policy

If you believe that this document breaches copyright please contact [librarypure@kcl.ac.uk](mailto:librarypure@kcl.ac.uk) providing details, and we will remove access to the work immediately and investigate your claim.

MEETING ABSTRACT

Open Access

# Pseudo-CT generation in brain MR-PET attenuation correction: comparison of several multi-atlas methods

Ines Merida<sup>1,2,3\*</sup>, Nicolas Costes<sup>2,4</sup>, Rolf Heckemann<sup>5</sup>, Alexander Hammers<sup>2,6</sup>

From PSMR 2015: 4th Conference on PET/MR and SPECT/MR  
La Biodola, Isola d'Elba, Italy. 17-21 May 2015

<sup>1</sup>Université de Lyon 1, INSERM, CNRS, Lyon Neuroscience Research Center, Lyon, France

Simultaneous MR-PET imaging opens new perspectives for understanding many aspects of brain function. However, accurate brain attenuation correction (AC) is required for absolute PET quantification. Radiodensity maps for AC are not readily available in PET-MR. Various strategies have been proposed to generate substitute attenuation maps. Recently, it has been shown that multiatlas techniques for AC outperform methods based on MR imaging sequences, such as Ultrashort-Echo-Time. We recently investigated several multiatlas methods to build a patient-specific pseudoCT, all based on multiatlas registration and label propagation, and introduced a novel maximum probability method (MaxProb). Here, a new database of 40 MR/CT image pairs (atlases) was used. Pairwise nonrigid registration of each atlas MR image to the target MR image yielded a geometric transformation that we used for propagating the atlas CT into the target's space. The synthetic CT was generated by voxelwise atlas selection and intensity averaging. We implemented a state-of-the-art algorithm along with two variants and MaxProb. A pseudoCT was generated for each subject of the database and assessed by comparison with the real CT in a leave-one-out scheme. Jaccard overlap indices were computed per tissue class (air, soft tissue, and bone). Voxel classification error was also assessed. The Jaccard indices and voxel classification error showed that MaxProb ( $8.31 \pm 1.12\%$ ) performed slightly better than Ref ( $8.65 \pm 0.93\%$ ). The influence of the number of atlases fused in pseudoCT generation was studied, showing that an optimal number improves accuracy. By adjusting the number of fused atlases, it was possible to significantly improve on Ref ( $8.46 \pm 0.96\%$ ). The methods now need evaluation on quantitative PET data to identify potential benefits in PET analysis.

## Authors' details

<sup>1</sup>Université de Lyon 1, INSERM, CNRS, Lyon Neuroscience Research Center, Lyon, France. <sup>2</sup>LILI-EQUIPEX – Lyon Integrated Life Imaging: hybrid MR-PET, Lyon, France. <sup>3</sup>Siemens Healthcare France SAS, Saint-Denis, Paris, France. <sup>4</sup>CERMEP-Imagerie du vivant, Lyon, France. <sup>5</sup>MedTech West at Sahlgrenska University Hospital, Gothenburg, Sweden. <sup>6</sup>Neurodis Foundation Lyon, Lyon, France.

Published: 18 May 2015

doi:10.1186/2197-7364-2-S1-A29

**Cite this article as:** Merida et al.: Pseudo-CT generation in brain MR-PET attenuation correction: comparison of several multi-atlas methods. *EJNMMI Physics* 2015 **2**(Suppl 1):A29.